



Production Test Report

Model:

Serial #:

Temperature Rating:

Manufacture date:

Room Temperature Test	Minimum	Maximum	Standard	Tolerance	Pass	Test Duration
	4998.0 mm	5001.0 mm	0.7 mm	±20 mm	YES	2Hrs

Cold Temperature Test	Minimum	Maximum	Standard	Tolerance	Pass	Test Duration
	859.0 mm	862.0 mm	0.4 mm	±10 mm	YES	4Hrs

Test Details:

- 1) Room temperature test with a target at a fixed distance of 5.00 Metres. 300 measurements are taken over a 2 hour period. The minimum, maximum and standard deviation
- 2) Cold test is conducted at a minimum of 5°C below the cold temperature rating with a target at a distance of ≈1.00 Metre. 600 measurements are taken over a 4 hour

Signature: 

 Renz Galaldez

 Edmonton, AB

 Canada

Date: December 11, 2018