



Production Test Report

Model: SR50A

 Temperature Rating: -45°C to +50°C

Serial #: 11055

Manufacture date: 1/9/2019

Room Temperature Test	Minimum	Maximum	Standard	Tolerance	Pass	Test Duration
	4994.0 mm	5000.0 mm	1.6 mm	±20 mm	YES	2Hrs

Cold Temperature Test	Minimum	Maximum	Standard	Tolerance	Pass	Test Duration
	859.0 mm	860.0 mm	0.4 mm	±10 mm	YES	4Hrs

Test Details:

- 1) Room temperature test with a target at a fixed distance of 5.00 Metres. 300 measurements are taken over a 2 hour period. The minimum, maximum and standard deviation
- 2) Cold test is conducted at a minimum of 5°C below the cold temperature rating with a target at a distance of ≈1.00 Metre. 600 measurements are taken over a 4 hour

Signature: Emir Muratovic
Date: January 14, 2019

Emir Muratovic
 Edmonton, AB
 Canada