



Production Test Report

Model:

Work Order#:

Serial #:

Temperature Rating:

Room Temperature	Minimum	Maximum	Standard	Tolerance	Pass	Test Duration
	5001.0 mm	5009.0 mm	1.8 mm	±20 mm	YES	2Hrs

Cold Temperature	Minimum	Maximum	Standard	Tolerance	Pass	Test Duration
	1008.0 mm	1011.0 mm	0.6 mm	±10 mm	YES	4Hrs

Test Details:

- 1) Room temperature test with a target at a fixed distance of 5.00 Metres. 300 measurements are taken over a 2 hour period. The minimum, maximum
- 2) Cold test is conducted at a minimum of 5°C below the cold temperature rating with a target at a distance of ≈1.00 Metre. 600 measurements are taken

Signature: 
Renz Geralddez
Edmonton, AB
Canada

Date: