

# Product certificate

Pages: 3  
 Release date: 15-06-2018

Product code **IR20-T2**  
 Product identification **serial number 4048**  
 Product type pyrgometer  
 Measurand longwave radiation

### Calibration result

Sensitivity  **$S_0 = 9.52 \times 10^{-6} \text{ V}/(\text{W}/\text{m}^2)$**   
 Calibration uncertainty  **$\pm 0.43 \times 10^{-6} \text{ V}/(\text{W}/\text{m}^2)$**

the number following the  $\pm$  symbol is the expanded uncertainty with a coverage factor  $k = 2$ , and defines an interval estimated to have a level of confidence of 95 percent

Temperature coefficients  
 **$a = -15.75 \times 10^{-6} \text{ }^\circ\text{C}^{-2}$**   
 **$b = 2.53 \times 10^{-3} \text{ }^\circ\text{C}^{-1}$**   
 **$c = 0.9556$**

Measurement function  $E = U/(S_0 \cdot (a \cdot T^2 + b \cdot T + c)) + \sigma \cdot (T + 273.15)^4$   
 with  $E$  irradiance in  $[\text{W}/\text{m}^2]$ ,  $U$  voltage output in  $[\text{V}]$ ,  $\sigma$  the Stefan-Boltzmann constant in  $[\text{W}/\text{m}^2/\text{K}^4]$ ,  $T$  the instrument body temperature in  $[\text{ }^\circ\text{C}]$

### Product specifications

1: resistance **10.2  $\Omega$**   
 2: insulation resistance **> 100 x 10<sup>6</sup>  $\Omega$**   
 3: cut-on wavelength (5 % transmission point)\* **4.4 x 10<sup>-6</sup> m**  
 4: spectral range (50 % transmission points)\* **4.7 to 40.0 x 10<sup>-6</sup> m**

\*values valid for this instrument, complete transmission curve available on request via info@hukseflux.com

**Table 0.1 connections**

PIN	WIRE	
2	Red	10 k $\Omega$ thermistor [+]
3	Pink	10 k $\Omega$ thermistor [+]
6	Blue	10 k $\Omega$ thermistor [-]
8	Grey	10 k $\Omega$ thermistor [-]
1	Brown	heater
4	Yellow	heater
9	Black	ground
7	White	signal [+]
5	Green	signal [-]

The 10 k $\Omega$  thermistor is a single four-wire thermistor measuring instrument body temperature.

Calibration procedure according to Hukseflux IR20C. Traceability of calibration is to the WISG (World Infrared Standard Group) operated at the Infrared Radiometry Section of the World Radiation Center in Davos, Switzerland.

Please consult the user manual for information on measurement uncertainty during actual use and for product set up, operation and maintenance instructions.

**Person authorising acceptance and release of product:**  
 M. Rietveld

**Date:**  
 15-06-2018

# Calibration certificate

Pages: 3  
Release date: 15-06-2018

Product code **IR20-T2**  
Product identification **serial number 4048**  
Product type pyrgeometer  
Measurand longwave radiation

### Calibration result

Sensitivity  **$S_0 = 9.52 \times 10^{-6} \text{ V}/(\text{W}/\text{m}^2)$**   
Calibration uncertainty  **$\pm 0.43 \times 10^{-6} \text{ V}/(\text{W}/\text{m}^2)$**

the number following the  $\pm$  symbol is the expanded uncertainty with a coverage factor  $k = 2$ , and defines an interval estimated to have a level of confidence of 95 percent

Reference conditions horizontal mounting, downward longwave irradiance, clear sky nights, 20 °C

### Measurement process

Metrological characteristic  $S_0$  in  $[\text{V}/(\text{W}/\text{m}^2)]$ ; sensitivity to downward longwave irradiance, with 180° field of view angle, valid for reference conditions  
Calibration method outdoor comparison to a reference pyrgeometer  
Measurement function  $E = U/(S_0 \cdot (a \cdot T^2 + b \cdot T + c)) + \sigma \cdot (T + 273.15)^4$  with E irradiance in  $[\text{W}/\text{m}^2]$ , U voltage output in [V],  $\sigma$  the Stefan-Boltzmann constant in  $[\text{W}/\text{m}^2/\text{K}^4]$ , T the instrument body temperature in [°C] and a, b, c temperature coefficients<sup>1</sup>  
Uncertainty of the method based on experience the expanded uncertainty is  $\pm 1.5\%$   
Measurement equipment Hukseflux Outdoor Test Facility  
Calibration conditions 2 clear sky night(s) between 06/06/2018 and 12-06-2018. Temperature between 11.5 °C and 15.9 °C. Net radiation between -76.4  $\text{W}/\text{m}^2$  and -70.0  $\text{W}/\text{m}^2$

### Metrological traceability

Calibration traceability to WISG (World Infrared Standard Group)  
Working standard pyrgeometer IR20, serial number 102  
Standard sensitivity  $S = 17.66 \times 10^{-6} \text{ V}/(\text{W}/\text{m}^2)$   
Calibration institute Physikalisch-Meteorologisch Observatorium Davos, World Radiation Center (PMOD/WRC)  
Uncertainty of standard  $\pm 3.4\%$  expanded uncertainty with a coverage factor  $k = 2$ , with respect to SI units

### Evaluation of the uncertainty of the calibration result

Uncertainty calculation the uncertainty is calculated as the square root of the sum of the squares of the reported uncertainties  
 $\sqrt{(1.5)^2 + (3.4)^2} = 4.5\%$

### Person performing calibration:

T. Meskers

### Date:

12-06-2018

<sup>1</sup> a, b and c are determined in the Hukseflux Temperature Response Characterisation, see separate page

# Temperature dependence

Pages: 3  
 Release date: 15-06-2018

Product code **IR20-T2**  
 Product identification **serial number 4048**  
 Product type pyrgometer  
 Measurand longwave radiation

**Characterisation result**

Temperature coefficients  
**a = -15.75 x 10<sup>-6</sup> °C<sup>-2</sup>**  
**b = 2.53 x 10<sup>-3</sup> °C<sup>-1</sup>**  
**c = 0.9556**

**Measurement process**

Characterised parameter dependence of sensitivity to temperature  
 Measurement function  $S(T) = S_0 \cdot (a \cdot T^2 + b \cdot T + c)$   
 with  $S(T)$  sensitivity in [V/(W/m<sup>2</sup>)] at an instrument body temperature  $T$ ,  $S_0$  sensitivity in [V/(W/m<sup>2</sup>)] at 20 °C instrument body temperature,  $T$  the instrument body temperature in [°C]  
 Measurement equipment Hukseflux Temperature Response Characterisation

**Table 0.2** temperature dependence test result

<b>TEMPERATURE DEPENDENCE TEST</b>									
T [°C]	-30	-20	-10	0	10	20	30	40	50
$(S(T) - S_0)/S_0$	-13.4%	-10.1%	-7.1%	-4.4%	-2.1%	0	+1.8%	+3.2%	+4.3%

**Person performing characterisation:**

L. Asaa

**Date:**

01-06-2018