

## Calibration certificate

 Pages 1  
 Release date: 26 FEB, 2019

Product code **SR30-D1**  
 Product identification **serial number 3186**  
 Product type pyranometer  
 Measurand hemispherical solar radiation

### Calibration result

Sensitivity  **$S = 10.37 \times 10^{-6} \text{ V}/(\text{W}/\text{m}^2)$**   
 Calibration uncertainty  **$\pm 0.10 \times 10^{-6} \text{ V}/(\text{W}/\text{m}^2)$**

the number following the  $\pm$  symbol is the expanded uncertainty with a coverage factor  $k = 2$ , and defines an interval estimated to have a level of confidence of 95 percent

Reference conditions 20 °C, normal incidence solar radiation, horizontal mounting, irradiance level 1000 W/m<sup>2</sup>

### Measurement process

Metrological characteristic S in [V/(W/m<sup>2</sup>)]: sensitivity to irradiance in the 300 to 3000 x 10<sup>-9</sup> m range, with 180° field of view angle, valid for reference conditions  
 Calibration method indoor calibration according to ISO 9847, type IIc  
 Measurement equipment Hukseflux Solar Radiation Calibration

### Metrological traceability

Calibration traceability to WRR (World Radiometric Reference)  
 Calibration hierarchy from WRR through ISO 9846 and ISO 9847  
 Working standard pyranometer type SR20, serial number 5039  
 Calibration institute PMOD World Radiation Center, Davos, Switzerland  
 Standard sensitivity 14.60 x 10<sup>-6</sup> V/(W/m<sup>2</sup>)

### Evaluation of the uncertainty of the calibration result

Uncertainty calculation the calibration uncertainty calculated as the square root of the sum of the squares of the calibration uncertainty of the working standard, the uncertainty of the method and the uncertainty due to deviations from the reference conditions is  $\pm 1.0 \%$ .

### Person performing calibration:

D.G. Dima

### Date:

22 FEB, 2019

### Person authorising calibration result of product:

H.E. Brouwer

### Date:

26 FEB, 2019

## Product certificate

 Pages: 4  
 Release date: 26 FEB, 2019

Product code	<b>SR30-D1</b>
Product identification	<b>serial number 3186</b>
Product type	pyranometer
Measurand	hemispherical solar radiation
Classification	secondary standard (ISO 9060), high quality (WMO-No. 8)

### Calibration result

Sensitivity	<b><math>S = 10.37 \times 10^{-6} \text{ V}/(\text{W}/\text{m}^2)</math></b>
Calibration uncertainty	<b><math>\pm 0.10 \times 10^{-6} \text{ V}/(\text{W}/\text{m}^2)</math></b>

the number following the  $\pm$  symbol is the expanded uncertainty with a coverage factor  $k = 2$ , and defines an interval estimated to have a level of confidence of 95 percent

### Product specifications and conformity

1:	ISO 9060 secondary standard	<b>verified</b>
2:	resistance	<b>18.4 <math>\Omega</math></b>
3:	insulation resistance	<b>&gt; 100 x 10<sup>6</sup> <math>\Omega</math></b>
4:	response time (95 %)	<b>2.8 s</b>
5:	temperature response	<b>verified</b>
6:	directional response	<b>verified</b>
7:	tilt measurement uncertainty	<b><math>\pm 1^\circ</math> (0 to 90 °)</b>

**Table 0.1** connections

PIN	WIRE	
1	Brown	VDC [+]
4	Black	VDC [-]
3	Blue	not connected
2	White	RS-485 B / B' [+]
5	Grey	RS-485 A / A' [-]
	Yellow	shield

Calibration procedure according to ISO 9847. Traceability of calibration is to the WRR (World Radiometric Reference) maintained at the World Radiation Center in Davos, Switzerland.

Please consult the user manual for set up, operation and maintenance instructions, and information on measurement uncertainty during actual use.

**Person authorising acceptance and release of product:**

H.E. Brouwer

**Date:**

26 FEB, 2019

## Directional response

Pages: 4  
Release date: 26 FEB, 2019

Product code	<b>SR30-D1</b>
Product identification	<b>serial number 3186</b>
Product type	pyranometer
Measurand	hemispherical solar radiation
Classification	secondary standard (ISO 9060), high quality (WMO-No. 8)

### Characterisation result

Directional response  $\leq \pm 5.8 \text{ W/m}^2$

### Measurement process

Characterised parameter dependence of sensitivity resulting from the direction of irradiance (a measure of the deviations from an ideal cosine response and its azimuthal variation)

Measurement functions  $C_{rel} = S(\theta)/(S(0) \cdot \cos(\theta) - 1) \cdot 100 \%$   
with  $C_{rel}$  the deviation from an ideal cosine response at zenith angle  $\theta$  in [%],  $S(\theta)$  the sensitivity to beam irradiance at zenith angle  $\theta$  in  $[\text{V}/(\text{W}/\text{m}^2)]$ ,  $S(0)$  the sensitivity to beam irradiance at normal incidence,  $\theta$  the incoming angle from zenith in  $[\circ]$

Measurement equipment  $C_{abs} = (S(\theta)/(S(0) \cdot \cos(\theta) - 1)) \cdot \cos(\theta) \cdot 1000$   
with  $C_{abs}$  the directional response as defined below in  $[\text{W}/\text{m}^2]$   
Hukseflux Directional Response Characterisation

### Conformity assessment

Definition of measurand The directional response is the error caused by assuming that the reported sensitivity is valid when measuring from any direction a beam whose normal incidence is  $1000 \text{ W/m}^2$

Acceptance interval ISO 9060 specifies a limit on the directional response for a secondary standard pyranometer of  $\pm 10 \text{ W/m}^2$

Conclusion Conformity verified

**Table 0.2** directional response test result

DIRECTIONAL RESPONSE TEST								
azimuth	North		East		South		West	
zenith	$C_{abs}$ [W/m <sup>2</sup> ]	$C_{rel}$ [%]	$C_{abs}$ [W/m <sup>2</sup> ]	$C_{rel}$ [%]	$C_{abs}$ [W/m <sup>2</sup> ]	$C_{rel}$ [%]	$C_{abs}$ [W/m <sup>2</sup> ]	$C_{rel}$ [%]
40 °	+2.0	+0.3	+3.9	+0.5	+5.8	+0.8	+4.1	+0.5
60 °	+0.4	+0.1	+3.9	+0.8	+5.6	+1.1	+1.3	+0.3
70 °	+0.1	+0.0	+3.3	+1.0	+4.2	+1.2	+0.0	+0.0
80 °	+0.0	+0.0	+3.9	+2.2	+3.8	+2.2	-0.2	-0.1

**Person performing characterisation:**

R. van den Dool

**Date:**

13 FEB, 2019

# Temperature response

Pages: 4  
 Release date: 26 FEB, 2019

Product code **SR30-D1**  
 Product identification **serial number 3186**  
 Product type pyranometer  
 Measurand hemispherical solar radiation  
 Classification secondary standard (ISO 9060), high quality (WMO-No. 8)

### Characterisation result

Temperature response **< ± 0.07 % (-30 to +50 °C)**  
 Temperature coefficients\*  
**a = -18.0490 x 10<sup>-6</sup> °C<sup>-2</sup>**  
**b = 7.3578 x 10<sup>-4</sup> °C<sup>-1</sup>**  
**c = 0.9925**

\* These temperature coefficients are applied internally in the instrument

### Measurement process

Characterised parameter dependence of sensitivity to ambient temperature  
 Measurement function  $S(T) = S_0 \cdot (a \cdot T^2 + b \cdot T + c)$   
 with  $S(T)$  sensitivity in [V/(W/m<sup>2</sup>)] at an instrument body temperature  
 $T$ ,  $S_0$  sensitivity at 20 °C instrument body temperature,  $T$  the  
 instrument body temperature in [°C],  $a$ ,  $b$  and  $c$  the temperature  
 coefficients determined from a second order polynomial fit  
 Measurement equipment Hukseflux Temperature Response Characterisation

### Conformity assessment

Definition of measurand\* Temperature response is the remaining percentage deviation in  
 sensitivity due to change in ambient temperature within a temperature  
 interval after the temperature coefficients are applied  
 Temperature interval -30 to +50 °C  
 Acceptance interval Hukseflux specifies a limit on the temperature response for a  
 SR30-D1 of ± 0.4 %  
 Conclusion Conformity verified

\* This is an adaptation of the definition in ISO 9060, which specifies a limit on the temperature response for a  
 secondary standard pyranometer of 2 % within a temperature interval of 50 K.

**Table 0.3** temperature dependence test result

TEMPERATURE DEPENDENCE TEST					
T [°C]	-30	-10	10	30	50
remaining deviation	+0.04 %	-0.07 %	+0.01 %	+0.06 %	-0.04 %

**Person performing characterisation:**  
 H.A. Kanij

**Date:**  
 14 FEB, 2019

# Tilt sensor characterisation

Pages: 4  
Release date: 26 FEB, 2019

Product code **SR30-D1**  
Product identification **serial number 3186**  
Product type pyranometer  
Measurand hemispherical solar radiation  
Classification secondary standard (ISO 9060), high quality (WMO-No. 8)

Characterisation result*	x-axis	y-axis	z-axis
gain	1.0086	0.9970	0.9956
offset	-184	6	-101
temperature coefficient $a_0$	$0.5295 \times 10^{-12}$	$-1.0831 \times 10^{-12}$	$-0.0062 \times 10^{-12}$
temperature coefficient $a_1$	-0.7779	-1.1574	-2.1771
temperature coefficient $a_2$	$-0.2331 \times 10^{-2}$	$-0.4768 \times 10^{-2}$	$0.3815 \times 10^{-2}$
temperature coefficient $a_3$	$0.4373 \times 10^{-4}$	$-1.0483 \times 10^{-4}$	$0.7278 \times 10^{-4}$

\* These gains, offsets and temperature coefficients are applied internally in the instrument

## Measurement process

Characterised parameters  
Measurement equation

tilt sensor gains and offsets  

$$\theta = 360/2\pi \cdot \text{atan}((x^2 + y^2)^{1/2}/z)$$

$$x, y, z = \text{gain}_{x,y,z} \cdot \text{raw}_{x,y,z} + \text{offset}_{x,y,z} + d_{x,y,z}(T)$$

$$d_{x,y,z}(T) = a_{0x,0y,0z} + a_{1x,1y,1z} \cdot T + a_{2x,2y,2z} \cdot T^2 + a_{3x,3y,3z} \cdot T^3$$
 with  $\theta$  the sensor tilt angle with respect to the horizontal in [°],  
 atan the arctangent function, x, y and z the corrected accelerometer counts,  $\text{gain}_{x,y,z}$  the tilt sensor gains,  $\text{raw}_{x,y,z}$  the raw accelerometer counts,  $\text{offset}_{x,y,z}$  the tilt sensor offsets,  $d_{x,y,z}(T)$  the correction for temperature dependence of the tilt measurement at an instrument body temperature T,  $a_0$ ,  $a_1$ ,  $a_2$  and  $a_3$  the temperature coefficients determined from a third order polynomial fit. Labels x, y and z refer to the three accelerometer axes.

Measurement process

Alignment with the bubble level is attained in horizontal position by introducing gains and offsets.  
 Gains and offsets are determined in horizontal position and at a tilt angle of 90 °. Temperature dependence of the tilt measurement is determined at a tilt angle of 90 ° between -30 and + 50 °C.

Measurement method

Hukseflux Tilt Sensor Characterisation

## Conformity assessment

Description of assessment  
Acceptance interval  
Conclusion

The tilt measurement uncertainty is verified in horizontal position and at a tilt angle of 90 °  
 The tilt measurement uncertainty is specified at  $\pm 1^\circ$  (0 to 90 °)  
 Conformity verified

**Person performing tilt sensor characterisation:**  
G.J. Halve

**Date:**  
25 FEB, 2019