

Notes

1. Analog
2. $R_3 = 50k \Omega$ 0.01%
3. MARC_program_101.CR1
4. Mobile Atmospheric Radiation Comparison
- 5.

The logger box mounts onto a surveyor tripod, door side facing up. The radiometers are stored in a Pelican case in the logger box. They mount to the outside of the door on plates. They are mounted with 2x M5 screws. The mounting does not need to be tight: just keep them from blowing away. Level them using the bubble level and leveling feet. Turn ON power at battery. Turn on fan but can be turned off to conserve power if needed. Data stored on logger: retrieve data using LoggerNet on the tablet

