

**Certificate of calibration**

Type: HFP01  
Serial Number: 003502

TIKST

Calibration Date: 31-03-2009

Sensitivity: 62.4  $\mu\text{V}/\text{Wm}^{-2}$

Remarks:

Performed by: R. Hartveld

Connections

White thermopile +  
Green thermopile -

Traceability of calibration is to the "guarded hot plate" of National Physical Laboratory (NPL) of the UK. Applicable standards are ISO 8302 (EN 12664:2001) and ASTM C177.

**Certificate of calibration**

Type: HFP01  
Serial Number: 003507

T1KST

Calibration Date: 02-04-2009

Sensitivity: 61.8  $\mu\text{V}/\text{Wm}^{-2}$

Remarks:

Performed by: R. Hartveld

Connections

White thermopile +  
Green thermopile -

Traceability of calibration is to the "guarded hot plate" of National Physical Laboratory (NPL) of the UK. Applicable standards are ISO 8302 (EN 12664:2001) and ASTM C177.