

LI-7500 CO₂/H₂O Analyzer

Calibration Certificate

Serial Number 75H-1694

Date: 10 Mar 2009

Technician JCF

CO₂ Calibration Values

A = 1.55023E2

B = 1.62293E4

C = 4.50913E7

D = -1.34814E10

E = 1.87114E12

XS = 0.0013

Z = -1.70000E-3

H₂O Calibration Values

A = 5.04086E3

B = 4.02623E6

C = -2.89168E8

XS = -0.0015

Z = 1.90000E-2

Pressure Calibration*

A0 = 10.761

A1 = 26.036

* Ver 3.0.1 and above

Zero/Span set on 11 Mar 2009

CO₂ Zero = 0.8603

CO₂ Span = 1.0025 (at 606 ppm)

H₂O Zero = 0.7945

H₂O Span = 0.9983 (at 12 C)



LI-COR, inc.
P.O. Box 4425
Lincoln, NE 68504 USA

Phone: 402-467-3576
FAX: 402-467-2819
Toll-free: 1-800-447-3576 (U.S. & Canada)

CO2 Measurements

	Type	Time	CO2(ppm)	CO2(mm3)	CO2Abs	CO2SD	CO2Slp	H2OAbs	TOven(C)	T(C)	Press	Pa(kPa)	Fs(l/m)	Cooler	AGC	Diag
1	CO2/75	19:15:52	0	0.000	0.00037	1.71E-5	-1.41E-5	-0.00276	16.60	15.73	97.88	97.99	0.501	1.4408	50.0	C
2	CO2/75	19:16:55	92.5	3.772	0.02413	1.54E-5	9.68E-6	-0.00283	16.60	15.73	97.89	97.99	0.500	1.4413	51.0	C
3	CO2/75	19:17:58	196.9	8.029	0.04715	1.55E-5	3.04E-5	-0.00287	16.60	15.76	97.90	98.00	0.501	1.4413	51.0	C
4	CO2/75	19:19:01	300.1	12.246	0.06686	1.48E-5	6.93E-6	-0.00291	16.60	15.74	97.96	98.01	0.500	1.4423	51.0	CH
5	CO2/75	19:20:04	398.4	16.254	0.08353	1.43E-5	9.63E-6	-0.00291	16.60	15.74	97.94	98.01	0.502	1.4433	52.0	C
6	CO2/75	19:21:07	505.5	20.624	0.09983	1.65E-5	5.71E-6	-0.00295	16.60	15.74	97.94	98.02	0.501	1.4423	52.0	CH
7	CO2/75	19:22:10	605.9	24.720	0.11369	1.43E-5	3.26E-5	-0.00298	16.60	15.74	97.94	98.02	0.502	1.4413	52.0	CH
8	CO2/75	19:23:13	747.3	30.486	0.13141	1.47E-5	4.32E-6	-0.00300	16.60	15.74	97.93	98.02	0.501	1.4413	52.0	CH
9	CO2/75	19:24:19	983.0	40.101	0.15738	1.45E-5	1.69E-5	-0.00303	16.60	15.74	97.93	98.02	0.500	1.4423	53.0	CH
10	CO2/75	19:25:30	1472	60.055	0.20162	1.73E-5	2.36E-5	-0.00308	16.60	15.74	97.94	98.02	0.501	1.4413	53.0	CH
11	CO2/75	19:26:39	2096	85.513	0.24668	1.55E-5	3.18E-5	-0.00312	16.60	15.74	97.94	98.02	0.502	1.4423	54.0	CH
12	CO2/75	19:27:51	2485	101.394	0.27065	1.77E-5	2.49E-5	-0.00315	16.60	15.74	97.95	98.03	0.500	1.4408	54.0	CH
13	CO2/75	19:29:04	3086	125.908	0.30348	1.80E-5	1.76E-5	-0.00319	16.60	15.76	97.95	98.04	0.501	1.4433	55.0	CH
14	CO2/75	22:52:58	0	0.000	-0.00013	1.89E-5	-7.01E-6	0.00291	31.90	30.75	98.44	98.48	0.501	1.7339	52.0	C
15	CO2/75	22:54:01	92.5	3.606	0.02270	1.83E-5	-1.01E-6	0.00277	31.90	30.77	98.46	98.48	0.501	1.7336	52.0	C
16	CO2/75	22:55:04	196.9	7.677	0.04500	2.23E-5	1.05E-5	0.00271	31.90	30.79	98.48	98.48	0.501	1.7326	53.0	CH
17	CO2/75	22:56:07	300.1	11.701	0.06422	1.73E-5	1.61E-5	0.00265	32.00	30.77	98.47	98.49	0.501	1.7346	53.0	C
18	CO2/75	22:57:22	398.4	15.531	0.08055	2.31E-5	1.32E-5	0.00264	31.90	30.79	98.46	98.49	0.501	1.7363	53.0	CH
19	CO2/75	22:58:25	505.5	19.706	0.09656	1.85E-5	2.23E-5	0.00260	31.90	30.79	98.46	98.50	0.501	1.7346	53.0	C
20	CO2/75	22:59:28	605.9	23.628	0.11024	2.36E-5	9.24E-6	0.00256	31.90	30.75	98.48	98.50	0.501	1.7323	54.0	C
21	CO2/75	23:00:31	747.3	29.149	0.12776	2.15E-5	1.80E-5	0.00253	31.90	30.77	98.51	98.50	0.500	1.7336	54.0	CH
22	CO2/75	23:01:37	983.0	38.340	0.15352	1.98E-5	2.10E-5	0.00250	31.90	30.79	98.51	98.50	0.501	1.7349	54.0	C
23	CO2/75	23:02:49	1472	57.391	0.19758	2.40E-5	8.97E-6	0.00245	31.90	30.75	98.46	98.51	0.500	1.7316	55.0	CH
24	CO2/75	23:04:04	2096	81.750	0.24265	2.02E-5	2.70E-5	0.00240	31.90	30.79	98.51	98.50	0.501	1.7326	56.0	C
25	CO2/75	23:05:11	2485	96.912	0.26671	1.72E-5	1.52E-5	0.00237	31.90	30.79	98.50	98.50	0.501	1.7356	56.0	C
26	CO2/75	23:06:14	3086	120.342	0.29974	2.25E-5	2.80E-6	0.00233	31.90	30.75	98.48	98.50	0.501	1.7366	57.0	CH
27	CO2/75	02:38:12	0	0.000	-0.00046	2.12E-5	-2.15E-5	0.00844	45.10	43.94	98.96	98.88	0.502	2.0214	54.0	C
28	CO2/75	02:39:18	92.5	3.473	0.02165	1.78E-5	-4.15E-6	0.00828	45.10	43.94	98.94	98.88	0.502	2.0222	55.0	C
29	CO2/75	02:40:25	196.9	7.391	0.04338	2.81E-5	2.38E-5	0.00822	45.10	43.99	98.92	98.88	0.501	2.0212	55.0	C
30	CO2/75	02:41:31	300.1	11.273	0.06221	2.08E-5	3.65E-5	0.00816	45.10	43.89	98.97	98.89	0.502	2.0222	55.0	C
31	CO2/75	02:42:34	398.4	14.969	0.07828	2.15E-5	1.63E-5	0.00814	45.20	43.99	99.02	98.89	0.501	2.0250	55.0	CH
32	CO2/75	02:43:43	505.5	18.987	0.09409	2.02E-5	9.59E-6	0.00809	45.10	43.92	98.97	98.90	0.501	2.0242	56.0	C
33	CO2/75	02:44:54	605.9	22.767	0.10764	1.97E-5	1.90E-5	0.00805	45.20	43.86	98.99	98.90	0.500	2.0202	56.0	C
34	CO2/75	02:45:57	747.3	28.073	0.12502	2.27E-5	3.10E-5	0.00802	45.10	43.92	98.98	98.91	0.500	2.0214	56.0	C
35	CO2/75	02:47:02	983.0	36.930	0.15063	1.81E-5	1.55E-5	0.00799	45.10	43.92	98.99	98.90	0.499	2.0212	57.0	CH
36	CO2/75	02:48:13	1472	55.285	0.19461	2.07E-5	3.53E-5	0.00794	45.20	43.89	98.95	98.91	0.500	2.0212	57.0	C
37	CO2/75	02:49:28	2096	78.737	0.23976	2.10E-5	1.94E-5	0.00788	45.20	43.89	98.97	98.91	0.500	2.0212	58.0	CH
38	CO2/75	02:50:31	2485	93.355	0.26392	1.77E-5	1.36E-5	0.00785	45.20	43.87	98.97	98.90	0.501	2.0222	58.0	CH
39	CO2/75	02:51:45	3086	115.910	0.29719	2.09E-5	1.18E-5	0.00781	45.10	43.87	98.95	98.90	0.500	2.0202	59.0	C

CO2(ppm) - CO2 concentration (tank value)

CO2(mm3) - Computed CO2 mole density (mmol/m3), based on mole fraction, temperature, and pressure.

CO2Abs - CO2 absorbance (unfiltered)

CO2SD - Standard deviation of CO2Abs (100 samples over 10 seconds).

H2OAbs - H2O absorbance (unfiltered)

TOven(C) - Oven temperature

T(C) - LI-7500's temperature measurement from calibration tube

Press - Atmospheric pressure measured by the LI-7500 (1% sensor)

Pa(kPa) - Atmospheric pressure measured by Ruska 6200

F(l/m) - Flow through calibration tube, liters/min.

Cooler - Detector cooling voltage.

AGC - Automatic gain control value (0-100%)

Diag - Diagnostics messages, plus 'C' for CO2 stability achieved, and 'H' for H2O stability achieved.

CO2 Computations

Num	ppm	abs/kPa	mmol/m3/kPa	Coeff's	Predicted	Error	%Error	Temp	Drift at 370 ppm	%C
1	0	0E0	0	1.55023E2	0	0	0.000	15C	-0.093	0.026
2	92.5	2.42811E-4	0.03853	1.62293E4	0.0392	0.00067	1.727	30C	0.285	
3	196.9	4.78011E-4	0.08201	4.50913E7	0.08208	0.00007	0.081	43C	0.627	
4	300.1	6.78998E-4	0.12501	-1.34814E10	0.12426	-0.00075	-0.598			
5	398.4	8.49406E-4	0.16596	1.87114E12	0.16483	-0.00113	-0.680			
6	505.5	1.0159E-3	0.21058		0.20918	-0.0014	-0.665			
7	605.9	1.15746E-3	0.2524		0.25079	-0.00161	-0.638			
8	747.3	1.33859E-3	0.3113		0.3095	-0.0018	-0.578			
9	983	1.60388E-3	0.40949		0.40708	-0.00241	-0.589			
10	1472	2.05559E-3	0.61318		0.60686	-0.00632	-1.030			
11	2096	2.51584E-3	0.87312		0.85926	-0.01385	-1.587			
12	2485	2.76039E-3	1.03516		1.01716	-0.018	-1.739			
13	3086	3.09568E-3	1.28543		1.26699	-0.01844	-1.434			
14	0	0E0	0		0	0	0.000			
15	92.5	2.31841E-4	0.03662		0.03734	0.00071	1.947			
16	196.9	4.58206E-4	0.07795		0.07822	0.00027	0.342			
17	300.1	6.53414E-4	0.11883		0.11857	-0.00026	-0.219			
18	398.4	8.19313E-4	0.15774		0.15732	-0.00042	-0.264			
19	505.5	9.81896E-4	0.20014		0.19973	-0.00042	-0.208			
20	605.9	1.12059E-3	0.23993		0.2396	-0.00033	-0.138			
21	747.3	1.29808E-3	0.2959		0.29582	-0.00008	-0.025			
22	983	1.55954E-3	0.3892		0.38978	0.00058	0.150			
23	1472	2.00776E-3	0.58289		0.58359	0.00071	0.122			
24	2096	2.4642E-3	0.82986		0.82819	-0.00167	-0.202			
25	2485	2.70868E-3	0.98388		0.98222	-0.00166	-0.169			
26	3086	3.04459E-3	1.22199		1.2261	0.0041	0.336			
27	0	0E0	0		0	0	0.000			
28	92.5	2.23366E-4	0.0351		0.03591	0.0008	2.292			
29	196.9	4.42983E-4	0.07472		0.07529	0.00057	0.767			
30	300.1	6.32931E-4	0.1139		0.11408	0.00018	0.155			
31	398.4	7.94827E-4	0.15117		0.15132	0.00015	0.101			
32	505.5	9.54901E-4	0.19185		0.19237	0.00052	0.272			
33	605.9	1.09153E-3	0.22999		0.23095	0.00096	0.416			
34	747.3	1.26715E-3	0.28362		0.2856	0.00197	0.695			
35	983	1.52561E-3	0.37307		0.37682	0.00376	1.007			
36	1472	1.97049E-3	0.55872		0.56582	0.00711	1.272			
37	2096	2.42608E-3	0.79556		0.80573	0.01017	1.278			
38	2485	2.67009E-3	0.94327		0.9567	0.01343	1.424			
39	3086	3.0067E-3	1.1714		1.19647	0.02507	2.140			

ppm - CO2 tank value (umol/mol)

abs/kPa - zero corrected CO2 absorptance divided by pressure

umolCO2/m3/kPa - CO2 mole density divided by pressure

Coeff's - computed calibration coeff's (fit 5th order poly to previous 2 columns)

Predicted - predicted CO2 (umol/m3/kPa)

% Error - (Predicted - actual)/actual * 100

At 370 - %Error at 370 ppm (based on curve fit of errors) for each temperature data set

%C - Estimated span drift with temperature at 370 ppm.

H2O Measurements

	Type	Time	H2O(C)	610kPa	H2O(ppt)	H2O(mm3)	CO2Abs	H2OAbs	H2OSD	H2OSlp	TOvent(C)	T(C)	Press	Pat(kPa)	Fs(l/m)	Cooler	AGC	Diag
1	H2O 75	19:34:26	-99.00	-0.22	0.00	0.000	0.00038	-0.00286	2.15E-5	2.73E-6	16.70	15.76	97.97	98.06	0.501	1.4423	50.0	CH
2	H2O 75	19:45:48	1.47	31.99	5.25	214.249	0.00043	0.03047	2.09E-5	-1.41E-5	16.70	15.78	97.98	98.07	0.250	1.4413	51.0	C
3	H2O 75	19:56:58	4.82	32.11	6.64	271.104	0.00043	0.03775	1.80E-5	-2.06E-5	16.70	15.76	98.02	98.11	0.250	1.4413	51.0	C
4	H2O 75	20:08:08	8.16	32.32	8.34	340.733	0.00044	0.04623	2.38E-5	2.04E-5	16.70	15.78	98.09	98.16	0.250	1.4428	51.0	C
5	H2O 75	20:19:18	11.51	32.42	10.43	425.989	0.00045	0.05565	2.15E-5	4.54E-6	16.70	15.81	98.07	98.18	0.250	1.4433	51.0	CH
6	H2O 75	20:30:28	15.68	32.73	13.65	557.768	0.00044	0.06880	1.70E-5	1.47E-5	16.70	15.79	98.11	98.21	0.250	1.4413	51.0	C
7	H2O 75	23:11:33	-99.00	-0.22	0.00	0.000	-0.00012	0.00264	2.40E-5	-6.95E-6	31.90	30.79	98.46	98.50	0.500	1.7363	52.0	CH
8	H2O 75	23:22:55	3.00	32.22	5.82	226.966	-0.00007	0.03750	2.43E-5	-1.41E-6	31.90	30.77	98.49	98.50	0.250	1.7333	53.0	CH
9	H2O 75	23:34:05	9.38	32.64	9.01	351.464	-0.00006	0.05277	1.85E-5	1.06E-5	32.00	30.81	98.53	98.51	0.250	1.7319	53.0	CH
10	H2O 75	23:45:15	15.75	32.98	13.65	532.444	-0.00003	0.07161	2.65E-5	7.91E-5	31.90	30.79	98.52	98.55	0.250	1.7343	53.0	C
11	H2O 75	23:56:25	22.14	33.49	20.27	791.179	0.00002	0.09478	2.05E-5	3.09E-5	31.90	30.81	98.59	98.58	0.250	1.7356	53.0	C
12	H2O 75	00:07:47	30.10	34.39	32.22	1257.356	0.00009	0.13054	6.85E-5	-7.45E-4	32.00	30.81	98.57	98.60	0.250	1.7346	54.0	C
13	H2O 75	02:57:10	-99.00	-0.21	0.00	0.000	-0.00044	0.00812	2.25E-5	-1.26E-5	45.20	43.86	98.97	98.90	0.500	2.0222	54.0	C
14	H2O 75	03:08:32	4.29	32.50	6.34	238.005	-0.00039	0.04418	2.08E-5	-1.46E-5	45.20	43.94	98.92	98.88	0.250	2.0212	55.0	C
15	H2O 75	03:19:58	13.32	33.13	11.63	436.770	-0.00037	0.06717	2.51E-5	9.77E-6	45.20	43.94	98.96	98.88	0.250	2.0202	55.0	CH
16	H2O 75	03:31:30	22.37	33.73	20.47	768.842	-0.00031	0.09805	2.70E-5	-2.43E-5	45.20	43.97	98.98	98.89	0.250	2.0224	56.0	C
17	H2O 75	03:43:05	31.40	34.72	34.55	1297.359	-0.00022	0.13840	1.01E-4	-7.55E-4	45.20	43.92	98.94	98.89	0.250	2.0212	56.0	C
18	H2O 75	03:55:12	42.70	36.62	63.11	2369.793	-0.00012	0.20485	3.44E-5	-2.99E-4	45.20	43.92	98.94	98.89	0.250	2.0212	57.0	C

H2O(C) - LI-610 Dewpoint generator set point

610kPa - Overpressure (kPa) in the LI-610

H2O(mm3) - Water mole density (mmol/m3) in the calibration tube. Includes overpressure correction.

H2OAbs - H2O absorbance (unfiltered)

H2OSD - Standard deviation of H2OAbs (100 samples over 10 seconds).

CO2Abs - CO2 absorbance (unfiltered)

TOvent(C) - Oven temperature

T(C) - LI-7000's temperature measurement

Pat(kPa) - Atmospheric pressure, measured by Ruska 6200

Press - Atmospheric pressure measured by LI-7500 (1 % sensor)

F(l/m) - Flow through calibration tube, liters/min.

Cooler - Detector cooler voltage

AGC - Automatic gain control value (0-100%)

Diag - LI-7500's diagnostic codes, plus 'C' for CO2 stability achieved, and 'H' for H2O stability achieved.

H2O Computations

Num	ppt	abs/kPa	mmol/m3/kPa	Coeffs	Predicted	Error	%Error	Temp	Drift at 10 ppt	%C
1	0	0E0	0	5.04086E3	0	0	0.000	15C	0.456	-0.014
2	5.2	3.39201E-4	2.18666	4.02623E6	2.16183	-0.02483	-1.136	30C	0.322	
3	6.6	4.13122E-4	2.7658	-2.89168E8	2.74926	-0.01655	-0.598	43C	0.057	
4	8.3	4.99032E-4	3.47368		3.48228	0.0086	0.248			
5	10.4	5.94913E-4	4.34372		4.36296	0.01924	0.443			
6	13.6	7.28322E-4	5.68513		5.69538	0.01025	0.180			
7	0	0E0	0		0	0	0.000			
8	5.8	3.54881E-4	2.30446		2.28305	-0.02141	-0.929			
9	9	5.10126E-4	3.56708		3.58083	0.01375	0.386			
10	13.7	7.01914E-4	5.40443		5.42191	0.01748	0.324			
11	20.3	9.37051E-4	8.02494		8.02092	-0.00402	-0.050			
12	32.2	1.30099E-3	12.75597		12.73605	-0.01992	-0.156			
13	0	0E0	0		0	0	0.000			
14	6.3	3.67521E-4	2.40604		2.3821	-0.02393	-0.995			
15	11.6	6.01591E-4	4.4136		4.42672	0.01312	0.297			
16	20.5	9.16005E-4	7.76765		7.77348	0.00583	0.075			
17	34.5	1.32754E-3	13.11258		13.11105	-0.00153	-0.012			
18	63.1	2.00465E-3	23.95182		23.95564	0.00382	0.016			

mmolH2O/mol - H2O concentration

abs/kPa - H2O absorbance divided by pressure

mmolH2O/m3/kPa - H2O mole density divided by pressure

Coeffs - computed calibration coeffs (fit 3rd order poly to previous 2 columns)







Predicted - predicted H2O (mmol/m3/kPa)

%Error - (Predicted - actual)/actual * 100







At 10 - Estimated % Error at 10 mmol/mol, based on a curve fit of the %Error values, for each temperature group

%C - Estimated span drift at 10 mmol/mol.



CO2 Calibration Plots



H2O Calibration Plots



CO2 Final Check Plots





File = L:\MICHAEL\CALS\75H\1694\20090311.fc

Type	Time	CO2(ppm)	CO2(mm3)	CO2Abs	CO2SD	CO2Slp	H2OAbs	CO2D	CO2DSD	H2OD	H2ODSD	TOven(C)	T(C)	Press	Pa(kPa)	Fs(l/m)	Cooler	AGC	Diag	
1	CO2.75	06:19:08	0	0.000	0.00000	1.70E-5	-4.74E-6	-0.00017	0.001	6.87E-4	-0.9	6.23E-2	26.90	25.86	98.97	99.02	0.502	1.6337	51.0	CH
2	CO2.75	06:20:23	196.9	7.846	0.04597	2.12E-5	2.22E-5	-0.00016	7.883	9.17E-4	-0.8	6.22E-2	26.90	25.85	99.01	99.02	0.502	1.6340	52.0	CH
3	CO2.75	06:21:38	300.1	11.959	0.06547	1.75E-5	1.02E-5	-0.00017	11.948	1.27E-3	-0.9	5.75E-2	26.90	25.85	99.01	99.03	0.501	1.6340	52.0	C
4	CO2.75	06:22:53	398.4	15.876	0.08201	1.98E-5	6.55E-6	-0.00014	15.849	1.58E-3	-0.7	6.40E-2	26.90	25.85	99.01	99.03	0.502	1.6357	53.0	CH
5	CO2.75	06:24:08	505.5	20.146	0.09821	2.34E-5	3.64E-7	-0.00016	20.120	1.54E-3	-0.8	6.56E-2	26.90	25.82	99.01	99.02	0.501	1.6337	53.0	CH
6	CO2.75	06:25:23	605.9	24.142	0.11204	1.63E-5	6.15E-6	-0.00017	24.139	2.12E-3	-0.9	6.71E-2	26.90	25.85	99.00	99.03	0.501	1.6350	53.0	C
7	CO2.75	06:26:38	747.3	29.776	0.12973	2.12E-5	-1.05E-5	-0.00017	29.808	2.79E-3	-0.8	7.01E-2	26.90	25.85	99.00	99.03	0.500	1.6347	53.0	CH
8	CO2.75	06:27:53	983.0	39.173	0.15572	1.51E-5	1.04E-5	-0.00016	39.256	3.61E-3	-0.8	6.64E-2	26.90	25.84	99.01	99.03	0.500	1.6340	54.0	C
9	CO2.75	06:29:08	1472	58.654	0.20009	2.02E-5	1.59E-5	-0.00014	58.659	6.91E-3	-0.7	6.76E-2	26.90	25.84	99.00	99.03	0.501	1.6330	54.0	CH
10	CO2.75	06:30:28	2096	83.537	0.24540	1.65E-5	1.27E-5	-0.00013	83.227	9.44E-3	-0.7	6.34E-2	26.90	25.80	99.01	99.04	0.501	1.6327	55.0	CH
11	CO2.75	06:31:43	2485	99.028	0.26954	1.80E-5	9.32E-6	-0.00012	98.625	1.40E-2	-0.6	6.80E-2	26.90	25.81	99.00	99.04	0.501	1.6337	55.0	C
12	CO2.75	06:33:23	3086	123.015	0.30269	1.96E-5	4.09E-5	-0.00011	122.983	1.68E-2	-0.6	6.54E-2	26.80	25.84	99.04	99.05	0.502	1.6350	56.0	CH

CO2(ppm) - CO2 concentration (tank value)
 CO2(mm3) - Actual value of CO2 mole density (mmol/m3)
 CO2Abs - CO2 absorbance (unfiltered)
 CO2D - Measured value of CO2 mole density (mmol/m3)
 CO2DSD - Standard deviation of CO2D (100 samples over 10 seconds).
 CErr - (CO2D - CO2(mm3))
 H2OD - Measured value of H2O mole density (mmol/m3)
 H2ODSD - Standard deviation of H2OD (100 samples over 10 seconds)
 T(C) - LI-7500's temperature measurement
 Press - LI-7500's pressure measurement (1% sensor)
 Pa(kPa) - Atmospheric pressure (measured by Ruskal 6200)
 Press - LI-7500's pressure measurement (kPa)
 F(l/m) - Flow through calibration tube, liters/min.
 AGC - Automatic gain control value (0-100%)
 Cooler - Detector cooler voltage
 Diag - LI-7500's diagnostic codes, plus 'C' indicates CO2 stability achieved, and 'H' indicates H2O stability achieved.

H2O Final Check Plots



File = L:\MICHAEL\CALS\75H\1694\20090311.fc

Type	Time	H2O(C)	610kPa	H2O(pppt)	H2O(mm3)	CO2Abs	H2OAbs	H2OSD	H2OSlp	CO2D	CO2DSD	H2OD	H2ODSD	TOven(C)	T(C)	Press	Pa(kPa)	Fs(l/m)	Cooler	AGC	Diag	
1	H2O 75	06:52:42	-99.00	-0.22	0.00	0.000	-0.00001	0.00000	4.29E-5	-8.01E-6	-0.002	7.48E-4	0.0	6.93E-2	26.90	25.86	99.09	99.11	0.500	1.6347	51.0	CH
2	H2O 75	07:10:12	11.82	32.38	10.55	422.154	-0.00002	0.05804	3.73E-5	-3.30E-5	-0.004	6.42E-4	422.7	1.18E-1	26.80	25.82	99.13	99.15	0.250	1.6340	52.0	C
3	H2O 75	07:21:40	19.83	32.95	17.57	701.087	-0.00004	0.08493	3.42E-5	2.47E-5	-0.006	6.52E-4	701.3	1.38E-1	26.90	25.81	99.13	99.19	0.250	1.6330	53.0	C
4	H2O 75	07:33:15	14.82	32.65	12.83	512.052	-0.00004	0.06726	3.71E-5	-5.61E-6	-0.006	6.68E-4	512.5	1.19E-1	26.90	25.81	99.15	99.21	0.250	1.6340	52.0	CH
5	H2O 75	07:45:00	9.84	32.33	9.27	370.033	-0.00004	0.05230	3.80E-5	1.47E-8	-0.005	7.27E-4	369.6	1.10E-1	26.90	25.85	99.18	99.24	0.250	1.6337	52.0	CH
6	H2O 75	07:56:57	4.84	32.04	6.60	263.507	-0.00003	0.03975	3.71E-5	-1.57E-5	-0.005	7.10E-4	262.1	1.12E-1	26.90	25.85	99.20	99.24	0.250	1.6340	52.0	C

H2O(C) - LI-610 Dewpoint generator set point

610kPa - Overpressure (kPa) in the LI-610

H2O(pppt) - True water concentration (mmol/mol)

H2O(mm3) - True water mole density (mmol/m3)

H2OAbs - H2O absorbance (unfiltered)

H2OD - Measured value of H2O mole density (mmol/m3)

H2ODSD - Standard deviation of H2OD (100 samples over 10 seconds)

HErr - (H2OD - H2O(mm3))

CO2D - Measured value of CO2 mole density (mmol/m3)

T(C) - LI-7500's temperature measurement

Press - LI-7500's pressure measurement (1% sensor)

Pa(kPa) - Atmospheric pressure (measured by Ruska 6200)

Press - LI-7500's pressure measurement (kPa)

F(l/m) - Flow through calibration tube, liters/min.

AGC - Automatic gain control value (0-100%)

Cooler - Detector cooler voltage

Diag - LI-7500's diagnostic codes, plus 'C' indicates CO2 stability achieved, and 'H' indicates H2O stability achieved.

Solar Response

Sun Angle (Deg)	Specification (ppm)	Typical (ppm)
15	10.0	< 1
25	9.2	< 1
35	8.0	< 1
45	6.0	< 1
55	4.0	< 1
65	3.2	< 1
75	3.0	< 1
85	3.0	< 1

This instrument has different filters from the original design. The above data is typical of this new design.

"Sun Angle" is the incidence angle on the source window.

"Specification" is the maximum allowable response.

"Typical" is the response by this instrument

('shaded' minus 'sunlit').

