

49i AND 49iPS DATA SHEET

MODEL # 49I-A3ZAA
 SERIAL # 0629719160

CUSTOMER # US DEPT. OF COMMERCE
 CO # 105740
 DATE 11/1/2006
 TECHNICIAN BILL A.

A FLOW 0.744
 B FLOW 0.750
 PRESSURE 748.5

TEMP
 BENCH 32.3
 LAMP 53.6
 03 LAMP 67.2

CALIBRATION
 03 BKG 0.00
 03 COEF 1.004

03	A	B	
<u>501</u>	<u>499</u>	<u>502</u>	1.5% * 500 = 7.5
<u>500</u>	<u>499</u>	<u>502</u>	507.5 A & B
<u>500</u>	<u>501</u>	<u>499</u>	492.5
<u>500</u>	<u>498</u>	<u>501</u>	
<u>500</u>	<u>496</u>	<u>505</u>	1% * 500 = 5
<u>500</u>	<u>500</u>	<u>500</u>	495 Average of A&B
<u>500</u>	<u>502</u>	<u>498</u>	505
<u>500</u>	<u>509</u>	<u>492</u>	
<u>500</u>	<u>504</u>	<u>496</u>	
<u>500</u>	<u>507</u>	<u>494</u>	
5001	5015	4989	
Average 500.1	501.5	498.9	
A/B Average	<u>500.2</u>		