



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Service
Report

RMA Number

74657A

Customer Information:

Company Woods Hole Oceanographic Institution **Date** 7/1/2013

Contact Robbie Laird

PO Number TBD

Serial Number 4537678-0122

Model Number SBE 45

Services Requested:

1. Evaluate/Repair Instrumentation.
2. Perform Routine Calibration Service.

Problems Found:

1. The anti-foulant devices appeared "dirty".

Services Performed:

1. Performed initial diagnostic evaluation.
2. Performed "Post Cruise" calibration of the temperature & conductivity sensors.
3. Installed NEW AF24173 Anti-foulant cylinder(s).
4. Performed internal inspection and O-ring replacement.
5. Performed hydrostatic pressure test.
6. Performed complete system check and full diagnostic evaluation.

Special Notes:

--