

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA
 Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3233
 CALIBRATION DATE: 27-Nov-12

SBE21 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -4.34600273e+000
 h = 5.12821545e-001
 i = -2.48087787e-004
 j = 3.88432900e-005
 CPcor = -9.5700e-008 (nominal)
 CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 8.75910177e-006
 b = 5.11962924e-001
 c = -4.34129445e+000
 d = -8.47180523e-005
 m = 4.4
 CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2.91225	0.00000	0.00000
1.0000	34.9016	2.98256	8.15855	2.98257	0.00001
4.5000	34.8810	3.29023	8.51675	3.29022	-0.00001
15.0000	34.8372	4.27391	9.57122	4.27390	-0.00000
18.5000	34.8275	4.61969	9.91479	4.61967	-0.00002
24.0000	34.8165	5.17865	10.44587	5.17868	0.00003
29.0000	34.8092	5.70129	10.91849	5.70130	0.00001
32.5000	34.8038	6.07406	11.24313	6.07404	-0.00001

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$\text{Conductivity} = (af^m + bf^2 + c + dt) / [10(1 + \epsilon p)] \text{ Siemens/meter}$$

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

