



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 3461
 CALIBRATION DATE: 12-Dec-18

SBE 4 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.00417448e+001
 h = 1.54557506e+000
 i = -2.45134692e-003
 j = 2.76679565e-004

CPcor = -9.5700e-008 (nominal)
 CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.55262	0.00000	0.00000
-1.0000	34.6346	2.79130	4.96409	2.79129	-0.00001
1.0000	34.6349	2.96193	5.07442	2.96194	0.00001
15.0000	34.6345	4.25166	5.84078	4.25166	-0.00001
18.5000	34.6328	4.59665	6.02916	4.59666	0.00002
29.0000	34.6274	5.67485	6.58281	5.67483	-0.00002
32.5000	34.6173	6.04520	6.76243	6.04522	0.00001

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;

Conductivity (S/m) = (g + h * f² + i * f³ + j * f⁴) / 10 (1 + δ * t + ε * p)

Residual (Siemens/meter) = instrument conductivity - bath conductivity

